


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-110A-P5*

S.N.: *8847*

Date: *MAY 11 1999*



a) Output signal amplitude:
0 TO ± 500 VOLTS

b) GAIN: *x 1 TO x 250*

c) Rise time: *≤ 35 ns*

d) Fall time: *≤ 35 ns.*

e) PRF: *DC TO 3 kHz*

f) Jitter, stability: *OK*

g) Prime power: *120/240V*

50-60 Hz

[Signature]

TR INPUT *2V/DIV*

PROT OUTPUT *≈ 200V/DIV*

200 ns/DIV.