


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3AH-C-N

S.N.: 8861

Date: JUNE 21 1999



- a) Output signal amplitude: 0 TO 450 Volts
- b) Pulse width: (75 ns)
- c) Rise time: 5 TO 100 ns
- d) Fall time: ≤ 1 ns
- e) PRF: ≤ 4 ns
- f) Jitter, stability: OK
- g) Prime power: 120/240 V
50-60 Hz

[Handwritten signature]