


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVL-3A.75-T

S.N.: 8864

Date: JUNE 21 1999



- a) Output signal amplitude: 0 to +300V (50Ω)
- b) Pulse width: 5 to 100ns
- c) Rise time: ≤ 600ps
- d) Fall time: ≤ 3ns
- e) PRF: 0 to 5kHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60MHz

70 dB ATTEN
320 V/DIV
5 ns/DIV
1 ns/DIV (RISE TIME)

PRF ≈ 3 kHz

M