


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVR4-3-B-PN-0A*

S.N.: *3265*

Date: *June 15 1989*



*(A) P<sub>out</sub>, 500V/div, R<sub>L</sub> = 10K.*



*(B) N<sub>out</sub>, 500V/div, R<sub>L</sub> = 10K.*

- a) Output signal amplitude:  
*0 to ±3000 V<sub>pk</sub>*
- b) Pulse width:  
*TO R<sub>L</sub> ≥ 10K.  
0.25 TO 2.5 μs*
- c) Rise time:  
*≤ 100 ns*
- d) Fall time:  
*≤ 100 ns*
- e) PRF: *0 TO 1K Hz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V  
50-60 Hz.*

*[Signature]*

