


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-C-BE03-R5-P-0PI-M

S.N.: 8880

Date: JUNE 7 1999



- a) Output signal amplitude:
0 TO +2 AMP ✓
- b) Pulse width: ✓
17 NS
- c) Rise time: ✓
≤ 2 NS
- d) Fall time: ✓
≤ 2 NS
- e) PRF: ✓
0 TO 20 KHz
- f) Jitter, stability:
OK ✓
- g) Prime power:

120/240V, 50-60 Hz ✓
[Handwritten signature]

MONITOR OUT TO 40 dB
100V/DIV, 2 AMP/DIV
5 NS/DIV

AVO-2T-BE02B3 CONNECTED
VIA A 8 METER CABLE
IN459A IN SERIES WITH
1 Ω USED AS TEST LOAD.
PRF = 20 KHz