


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1-C-P-M-OT-U01

S.N.: 8903

Date: JULY 15 1999



- a) Output signal amplitude:
0 TO +5 V (50Ω)
- b) Pulse width:
0.25 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 125 ns
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power:
120/240V
50-60Hz

40 db ATTEN

0.5 VOLTS/DIV

PRF = 300 kHz

TDP 500ps/DIV

POT 2 ns/DIV

120/240V
50-60Hz
AM