


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-HUB-GE03*

S.N.: *8918*

Date: *Jul 19 1999*



*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 300 kHz*

- a) Output signal amplitude: ✓  
*± 20 Vrms (± 50V)*
  - b) Pulse width: ✓  
*≤ 350 ps*
  - c) Rise time: ✓  
*≤ 100 ps*
  - d) Fall time: ✓  
*≤ 200 ps*
  - e) PRF: ✓  
*0.90 1.0 MHz*
  - f) Jitter, stability: ✓  
*OK*
  - g) Prime power:
    - a) *+15V, 35 mA MAX*
    - b) *-15V, -35 mA MAX*
    - c) *+135V, 4 mA MAX*
- [Signature]*