


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-6C-B7

S.N.: 3927

Date: JUL 13 1999



- a) Output signal amplitude:
0 TO +5 AMPS
- b) Pulse width:
50 NS TO 5.0 US
- c) Rise time: 1% MAX DUTY CYCLE
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: 0 TO 10 KHz
1% MAX DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power: 120/240 V
50-60 Hz

PEARSON 410 OUTPUT
≈ 5 AMP/DIV.

TOP 100 NS/DIV

BOI 500 NS/DIV.

RG