


PULSE GENERATOR
PERFORMANCE CHECK

Model: *M-110C-PS*
S.N.: *8940*
Date: *JULY 16 1999*



a) Output signal amplitude:
*0 TO ±100V (TO RL
≥10K)*

GAIN: *0 TO X50*

e) PRF: *DC TO 50KHz*

f) Jitter, stability: *OK*

g) Prime power: *120/200V
50-60 Hz.*

TOP: IN, 2V/DIV

*BOT: out, 50V/DIV
RL = 10K.*