


PULSE GENERATOR
PERFORMANCE CHECK

Model: MUNN-1-PS
 S.N.: 8945
 Date: SEPT 3 1999



- a) Output signal amplitude: 0 TO 5 VOLTS
- b) Pulse width: (50 ns)
1 TO 30 ns
- c) Rise time: ≤ 100 ps
- d) Fall time: ≤ 100 ps
- e) PRF: ≤ 100 /s
25 TO 250 MHz
- f) Jitter, stability: OK

g) Prime power: 120/240 V

40 dB ATTN.: 5 V/DIV
TOP 2 ns/DIV
MID 200 ps/DIV (RISE TIME)
BOT 200 ps/DIV (FALL TIME)
PRF ≈ 100 MHz.

50-60 MHz
[Signature]