


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M-151B-C*

S.N.: *8947*

Date: *JUNE 10, 1999*



- a) Output signal amplitude: *0 TO ± 200V ✓*
- b) Pulse width: *( $R_L \geq 50\Omega$ )  
SINE, SQUARE, TRIANGLE ✓*
- c) Rise time: *+DC OFFSET  
≤ 100ns ✓*
- d) Fall time: *≤ 100ns ✓*
- e) PRF: *0 TO 100kHz ✓*
- f) Jitter, stability: *✓*
- g) Prime power: *120/240V ✓  
50 to 60kHz ✓*

*10 us / DIV  
50 V / DIV  
 $R_L = 50k$*

*[Signature]*