


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-HVI-C-P-PW-M-05*

S.N.: *8969*

Date: *JUN 20 1979*



- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100NS
- c) Rise time:
≤ 250ps
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

50.46 ATTEN

32 VOLTS/DIV.

TOP: 10 ns/DIV.

*BOT: 200ps/DIV
(RISE TIME)*

PRF ≈ 10 KHz

[Signature]