


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-2-P*

S.N.: *8970*

Date: *SEPT 7 1999*



- a) Output signal amplitude: *0 to +120V*
- b) Pulse width: *(50%)*  
*≤ 2 ns*
- c) Rise time: *≤ 1 ns*
- d) Fall time: *≤ 1 ns*
- e) PRF: *0 to 50 kHz*

- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS*  
*250 mA MAX.*

*60 db ATN*  
*100 VOLTS/DIV*  
*2 NS/DIV*  
*PRF ≈ 30 kHz*