


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14F-3*

S.N.: *8972*

Date: *SEPT 8 1999*



- a) Output signal amplitude: *10V*
- b) Pulse width: *130 ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *500 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *500 MHz*
0.5 WATT
(15 VOLT P-P).

40 dB ATTEN
10 VOLTS/DIV

* *15 VOLT SPECIFICATION*
IN TABLE ON PAGE 78
OF CAT 10 IS A MISPRINT.