


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PVMM-2-PS-TN-N'*

S.N.: *8982*

Date: *SEPT 13 1999*



- a) Output signal amplitude: *0 to -5V (to 50mV)*
- b) Pulse width: *1 to 10ns*
- c) Rise time: *300ps to 2.0ns*
- d) Fall time: *300ps to 2.0ns*
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability: *OK.*
- g) Prime power: *100 ± 10 volts*

*40dB ATTN
0.5 VOLTS/DIV
TR SWITCH IN
5 POSITIONS.*

50-60Hz
[Signature]