


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-6C-B-P*
S.N.: *8995*
Date: *NOV 1 1999*

- a) Output signal amplitude: *0 TO + 5 AMPS*
- b) Pulse width: *50 NS TO 5.0US*
- c) Rise time: *(19. MAX DUTY CYCLE)
≤ 10 NS*
- d) Fall time: *(20-30%)
≤ 10 NS (20-80%)*
- e) PRF: *0 TO 10K Hz*
- f) Jitter, stability: *(19. MAX DUTY CYCLE)
OK*
- g) Prime power: *120 / 240V
50 - 60 Hz*

AMERICAN LASER MODEL 711
CURRENT PROBE OUTPUT.
INSO19 TEST LOAD.
= 5 AMPS / DIV.



MAINFRAME OUTPUT TO 50Ω

