


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1-P-M-EW*

S.N.: *10014*

Date: *OCT 26 2001*



- a) Output signal amplitude:  
*0 TO +10V (to 50  $\Omega$ )*
- b) Pulse width:  
*0.2 TO 40 NS*
- c) Rise time:  
 *$\leq 100$  ps*
- d) Fall time:  
 *$\leq 135$  ps*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VDC*  
*250 mA MAX.*

*40 db. ATTN*  
*10 VERT/DIV*  
*TOP 1 NS/DIV*  
*BOT 100 ps/DIV*  
*(RISE TIME)*

*PRF = 100 kHz*