


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVH-M2-N*
S.N.: *10017*
Date: *DEC 10 2001*



- a) Output signal amplitude:
≥ -60V (70 50n)
- b) Pulse width:
≤ 2 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

+135 ± 5 VDC
5 mA (max)

50 dB ATTN

≈ 32 V/DIV

100 kHz PRF