


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-M2-N*

S.N.: *10019*

Date: *DEC 10 2001*



- a) Output signal amplitude:
71 - 60V (to 50.5)
- b) Pulse width:
≤ 2 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 to 1 MHz.*
- f) Jitter, stability:
OK
- g) Prime power:

50 Ω 16 AATTN
≈ 32 V / DIV
100 kHz PRF

+135 ± 5 VDC
5 mA (MAX)