

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR4-1C-PN-PWT

S.N.: 10053



Date: Dec 7 2001

- a) Output signal amplitude:
0 TO ± 1000 VOLT
- b) Pulse width: (TO $R_L \geq 200 \Omega$)
0.2 μ S TO 5.0 μ S
- c) Rise time:
 ≤ 50 NS
- d) Fall time:
 ≤ 50 NS
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
170/240 V
50-60 Hz.



Q

10053
P out
 $R_L \approx 1000 \Omega$
PRF = 1 KHz




ⓑ

10053

N out.

$R_C \approx 10K$

PRF = 1K



LOAD 13 BE ⓐ