


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-PS-P-PN*

S.N.: *10070*

Date: *DEC 20 2001*



- a) Output signal amplitude:
0 TO ±15 V (P.50Ω)
- b) Pulse width:
0.2 TO 2.0 N
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

40db ATTEN.: 10 V/DIV
PRF ≈ 10 MHz.

50-60 MHz.

[Signature]