


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-4C-C-PN-PA-05*

S.N.: *10093 (MOD 1)*

Date: *MAR 6 2002*



- a) Output signal amplitude:
+150 TO +1000V (70.50n)
- b) Pulse width:
≤ 8 NS
- c) Rise time:
≤ 2.5 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:

70 dB ATTEN

320V/DIV

PRF = 5 KHz

120/240V


50-60 Hz

PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVG-4C-PS-PN-6A-05*

S.N.: *10093*

Date: *DEC 21 2001*



- a) Output signal amplitude:
+150 TO +1000 V (TO 50W)
- b) Pulse width:
≤ 8 NS
- c) Rise time:
≤ 2.5 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 to 100 Hz

70 db ATTEN
320 VOLTS/DIV.
PRF = 5 KHz.