


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-M-14V3-C-P-DN-BA-EW-EU-ELL

S.N.: 10137

Date: MAR 14 2002



- a) Output signal amplitude: 0 TO ± 40 VOLTS (90)
- b) Pulse width: 500ns
- c) Rise time: 0.4 TO 2.0nS
- d) Fall time: ≤ 150ps
- e) PRF: ≤ 250ps
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60MHz

50 dB ATTEN
≈ 32 VOLTS/PW
PW max + PW min
PRF = 100 kHz