


PULSE GENERATOR  
PERFORMANCE CHECK

Model: ANI-V-HV3-C-PN

S.N.: 10209

Date: JUNE 21 2002



- a) Output signal amplitude: 0 TO  $\pm 100$  VOLTS
- b) Pulse width: 5 NS/DIV TO 50 NS
- c) Rise time: 4 TO 100 NS
- d) Fall time:  $\leq 200$  ps (20-80 ps)
- e) PRF:  $\leq 4$  MS
- f) Jitter, stability: OK
- g) Prime power: 0 TO 5 KHz

60 dB ATTN  
- 100 VOLTS/DIV  
PRF = 5 KHz

120/240 VOLTS  
50-60 Hz