


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-3A-C-P-0A*

S.N.: *10214*

Date: *JUNE 24 2002*



- a) Output signal amplitude:
0 to +300 V (to 50Ω)
- b) Pulse width:
5 to 100 ns
- c) Rise time:
≤ 600 ps
- d) Fall time:
≤ 3 ns
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
5 to 60 kHz

60 dB ATTEN

100 V/DIV

TRIG 10 ns/DIV

ROUT 1 ns/DIV (RISE TIME)

PRF = 2.0 kHz

[Signature]