


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-1-C-P*

S.N.: *10264*

Date: *JUL 17 2002*



- a) Output signal amplitude:
0 to +10V (70.50ns)
- b) Pulse width:
≤ 200 ps.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 Volts
50-60 Hz.

40 dB ATTN
10 V/DIV
TOP 1 ns/DIV
POT 200 ps/DIV

PRF ≈ 12 MHz.

[Signature]