


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-N-15-C-P*

S.N.: *10268*

Date: *MAY 29 2002*



- a) Output signal amplitude:  
*0 TO +5V (TO 50V)*
- b) Pulse width:  
*0.2 TO 4.0NS*
- c) Rise time:  
*≤ 60 ps*
- d) Fall time:  
*≤ 135ps (AT PW MIN)*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 dB ATTN*  
*- 10 VOLTS/DIV*

*TOP 1 NS/DIV*

*BOT 100ps/DIV (RISE TIME)*

*120/240V*  
*50-60 Hz.*