


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PMR-53-C-P-QTKA*

S.N.: *10275*

Date: *JULY 30 2002*



- a) Output signal amplitude:  
*0 to +40V (to 50Ω)*
- b) Pulse width:  
*1.5 to 10.0 ns*
- c) Rise time:  
*0.5 ns (20-80%)*
- d) Fall time:  
*≤ 1.0 ns*
- e) PRF:  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB ATTN*

*32 VOLTS/DIV*

*2 NS/DIV*

*PRF = 100 KCF*

*120/240V*

*50-60 Hz*