


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AMP-N-2-P  
S.N.: 10315  
Date: JUN 18 2002



- a) Output signal amplitude:  
0 TO +10 VOLTS (TO 50V)
- b) Pulse width:  
2 TO 50NS
- c) Rise time:  
≤ 100ps
- d) Fall time:  
≤ 135ps
- e) PRF:  
0 TO 50 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VDC  
400 mA (MAX)

DR 5 NS/IN  
AMP 200ps/DIV (RISE TIME)  
50T 200ps/DIV (FALL TIME)  
40 dB ATTN (10 VOLTS/DIV)  
PRF = 50KHz

[Signature]