


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVP-25-PS-P

S.N.: 10355

Date: SEPT 13 2002



- a) Output signal amplitude: ✓  
0 TO +2 VOLTS
- b) Pulse width: ✓  
0.2 TO 4.0 NS
- c) Rise time: ✓  
≤ 40 PS
- d) Fall time: ✓  
≤ 135 PS
- e) PRF: ✓  
0 TO 1.0 MHz
- f) Jitter, stability: ✓  
OK
- g) Prime power: ✓  
120/240V  
50-60Hz.

20 dB ATTEN  
∴ 2 VOLTS / DIV  
PRF = 1.0 MHz

$$T_{RT} = \sqrt{T_{R25}^2 + T_{SCOPE}^2 + T_{ATTEN}^2}$$

$$\approx 50 \text{ PS (20-80\%)}$$

$$T_{SCOPE} = 25 \text{ PS}$$

18 GHz ATTEN + CABLES.

∴ CONCL THAT  $T_{R25} \leq 40 \text{ PS}$