


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PUL-3A.H-C-P*

S.N.: *10369*

Date: *Oct 21 2002*



- a) Output signal amplitude:
0 TO +450V
- b) Pulse width: *(TO 50ns)*
5 TO 100ns
- c) Rise time:
≤ 1ns
- d) Fall time:
≤ 4ns
- e) PRF: *0 TO 5KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240V
50 - 60 Hz

60dB ATTN

- 100 VOLTS/DIV

TOP 10 NS/DIV

BOI 2 NS/DIV (RISE TIME)

PRF = 2.5 KHz

120 / 240V
50 - 60 Hz
[Signature]