


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV6-4C-C-P*  
S.N.: *10400*  
Date: *OCT 30 2002*



- a) Output signal amplitude:  
*0 TO +1000 V (70 50n)*
- b) Pulse width:  
*≤ 8ns*
- c) Rise time:  
*≤ 2ns*
- d) Fall time:  
*≤ 4ns*
- e) PRF: *0 TO 10 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*170 / 200 V*  
*50 - 60 Hz*

*70 dB ATTN.*  
*≈ 320 VOLTS / DIV.*  
*PRF = 5 kHz.*