


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-1-C-P*
S.N.: *10415*
Date: *AUG 27 2002*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

40 dB ATTN
5 VOLTS/DIV
TOP 1 NS/DIV
BOT 200 ps/DIV