


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P-DN-BA-EW-00*

S.N.: *10423*

Date: *NOV 7 2002*



- a) Output signal amplitude:
0.70 ± 5 VOLTS (70.50V)
- b) Pulse width:
0.270 (6N)
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0.70 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

40 LB ATTEN
.5 VOLTS/DIV
TOP 1NS/DIV
BOI 100ps/DIV
(RISE TIME)
PRF = 12 MHz