


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-143-C-N-PN-EA ED-EW-M*

S.N.: *10431*

Date: *NOV 19 2002*



- a) Output signal amplitude:
0 TO 740 VOLTS
- b) Pulse width: *(TO 50ns)*
0.4 TO 2.0NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF: *0 TO 1 MHz.*
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz.

[Signature]

50 dB ATTN

32 VOLTS/DIV

PRF = 100 KHz

TOP PW AT MIN.

BOT PW AT MAX.