


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-2-P*
S.N.: *10440*
Date: *NOV 13 2002*



- a) Output signal amplitude: *0 to +120 V*
- b) Pulse width: *(to 50ns) ≤ 2ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

60 dB ATTN
- 100 VOLTS / DIV
PRF = 10 KHz

+15 VDC
300 mA (MAX)