


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MAP-10-P-M.*

S.N.: *10446*

Date: *DEC 16 2002*



40 dB ATTEN., 10V/DIV

- a) Output signal amplitude: *0 TO +10V (TO 50N)*
- b) Pulse width: *0.2 TO 100 NS*
- c) Rise time: *≤ 100 ps.*
- d) Fall time: *≤ 135 ps*
- e) PRF: *0 TO 1.0 MHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz.

① A RANGE *1.0 NS/DIV*

② A RANGE *100 ps/DIV (RISE TIME)*

③ B RANGE *10 NS/DIV*

④ B RANGE *100 ps/DIV RISE TIME*

[Signature]