


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AU-4V3-C-P-6A-50-EW-M*

S.N.: *10482*

Date: *JAN 27 2003*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width: *(TO 50N)*
0.4 TO 2.0NS
- c) Rise time:
≤ 150ps
- d) Fall time:
≤ 250ps
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

50 dB ATTN
32 VOLTS/DIV
PRF = 100 KHz
TOP MARK PW
BOT MARK PW

120/240V
50-60 Hz
[Signature]