


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-14V3-C-N*

S.N.: *10495*

Date: *FEB 10 2003*



- a) Output signal amplitude: *0 to -40V (70.50Ω)*
- b) Pulse width: *0.4 to 2.0 ns*
- c) Rise time: *≤ 150 ps*
- d) Fall time: *≤ 25 ps*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60 Hz

50 dB ATTN
-32 VOLTS/DIV

TOP PW NORMAL MIN
BOT PW NORMAL MIN
PRF = 100 kHz