

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-C-N-KOPA-DA*

S.N.: *10529*

Date: *FEB 26 2003*

- a) Output signal amplitude:
0 to -300V (to $R_L \geq 50\Omega$)
- b) Pulse width:
100 ns to 1.0 μ s
- c) Rise time:
 ≤ 15 ns
- d) Fall time:
 ≤ 15 ns
- e) PRF:
0 to 1 kHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]

10529

BOTH outputs $R_L = 50\Omega$
 $P_{RF} = 1\text{KW}$

