


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMM-2-P-GW*

S.N.: *10534*

Date: *JAN 23 2003*



- a) Output signal amplitude:  
*0 TO +5 VOLT (TO 50N)*
- b) Pulse width:  
*1 TO 10 NS*
- c) Rise time:  
*≤ 300ps*
- d) Fall time:  
*≤ 600ps.*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 dB ATTN*

*5 VOLT/DIV*

*PRF = 10 MHz.*

*+24 VDC,*

*500 mA (LIMIT).*