


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV4-5B-7*

S.N.: *10548*

Date: *JAN 06 2004*



- a) Output signal amplitude:
0 to +30V (TDS02)
- b) Pulse width:
≤ 350 μs
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTN
10 V/div / DIV
PRF = 100 kHz

+15 VDC
200 mA MAX.