


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-AV1-CA5-M1-T*

S.N.: *10560*

Date: *FEB 14 2003*



- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
≤ 1.0 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 200 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:

+ 15 VDC

250 mA MAX

60 dB ATTN
100 VOLTS/DIV
PRF = 100 KHz

[Signature]