


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-MPT-EBBIE*

S.N.: *10597*

Date: *MAR 26 2003*



- a) Output signal amplitude: ✓  
*≥ +10 V<sub>OH</sub> (70%<sub>±3</sub>)*
- b) Pulse width:  
*1.5 NS FWHM*
- c) Rise time: ✓  
*≤ 100 ps*
- d) Fall time:  
*≈ 3 NS ✓*
- e) PRF:  
*0 TO 2 MHz.*
- f) Jitter, stability:  
*OK ✓*
- g) Prime power:
  - a) *+15 VDC*  
*160 mA*
  - b) *-5 VDC*  
*10 mA*

*40 db ATTEN*  
*10 V<sub>OH</sub> / DIV*  
*TOP: 1 NS / DIV*  
*BOT: 100 ps / DIV*  
*PRF = 2.0 MHz*

*[Signature]*