


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-N*

S.N.: *10652*

Date: *JULY 22 2003*



- a) Output signal amplitude:  
*0.70 - 5.00V (70.50Ω)*
- b) Pulse width:  
*0.2 TO 6.0 NS*
- c) Rise time:  
*≤ 100 ps.*
- d) Fall time:  
*≤ 135 ps*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40dB ATTEN*

*5 VOLTS/DIV*

*TOP 1.0 NS/DIV*

*BOE 100ps/DIV  
(RISETIME)*

*PRF = 12 MHz.*

*120/240V*

*50 - 60 Hz*

*[Signature]*