


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV124-C-N-AT-PWT*

S.N.: *10671*

Date: *AUG 8 2003*



- a) Output signal amplitude:
0 TO 200 V (TO 50 W)
- b) Pulse width:
2 TO 200 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 20 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz

[Signature]

60 dB ATTN
∴ 100 VOLTS/DIV

TOP *2-10 NS RANGE,*
5 NS/DIV

BOT *10-200 NS RANGE*
10 NS/DIV
PRF = 20 KHz