


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMH-5-C-P*  
S.N.: *10682*  
Date: *SEPT 4 2003*



- a) Output signal amplitude:  
*0 TO +100 V (70.50V)*
- b) Pulse width:  
*2 TO 4 NS*
- c) Rise time:  
*≈ 1 NS*
- d) Fall time:  
*≈ 2 NS*
- e) PRF:  
*0 TO 1 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/2500 V*  
*50-60 MHz.*

*60dB ATTN*  
*100 V 283/DIV*  
*PRF = 100 KHz.*

*[Signature]*