


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM4-1-C-P-M*

S.N.: *10722*

Date: *OCT 14 2003*



- a) Output signal amplitude:  
*0 TO +10V (70.50n)*
- b) Pulse width:  
*≤ 200ns*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *1 TO 25 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*\* 50-60Hz*

*[Signature]*