


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANP-A-HV3-P-05*

S.N.: *10726*

Date: *OCT 28 2003*



- a) Output signal amplitude:  
*0 TO +40 (to 50V)*
- b) Pulse width:  
*0.4 TO 2 NS*
- c) Rise time:  
*≤ 150ps*
- d) Fall time:  
*≤ 250ps*
- e) PRF:  
*0 TO 1.0 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50dB ATTEN - : 32 V/DIV*

*PRF = 100 KHz.*

*TOP 1 NS/DIV*

*Bot 100ps/DIV (RISE TIME)*

*+15V DC*

*250 mA MAX.*

*[Signature]*