


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-N-HV3-P-05*

S.N.: *10727*

Date: *OCT 28 2003*



- a) Output signal amplitude:
0 TO +40 (TO 50V)
- b) Pulse width:
0.4 TO 2 NS
- c) Rise time:
≤ 150ps
- d) Fall time:
≤ 250ps
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
+15V DC
250 mA MAX.

50dB ATTEN - : 32 V/DIV

PRF = 100 KHz.

TOP PW MAX

BOT PW MIN

1. NS / DIV.